



VITROMED
Germany

VITROMED GmbH
Raiffeisenstraße 15a
40764 Langenfeld
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 002M0901
PRODUCT NAME: V-SPERM WASH
REFERENCE: V-SWM-02
EXPIRY DATE: 28.02.2021

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7,39
OSMOLALITY	257-273 mOsm/kg	266
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED

PRODUCT RELEASED BY Lela Abazi ON 04.06.2020

Lela Abazi
