

CERTIFICATE OF ANALYSIS

LOT NUMBER: 003M3602
PRODUCT NAME: V-HYADASE
REFERENCE: V-HYA
EXPIRY DATE: 31-Mar-2021

TEST	SPECIFICATION	RESULT
pH	7,20 – 7,40	7,34
OSMOLALITY	257 – 273 mOsm/kg	268
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,00 EU/ml	0,126
MEA 1-cell to expanded blastocyst 96 h	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED