



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19589VD40, 19629VD35, 19630VD75, 19631VD20, 19632VD27, 19633VD30

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-140, V-DEN-135, V-DEN-175, V-DEN-200, V-DEN-275, V-DEN-300

EXPIRY DATE: 28.01.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED