

CERTIFICATE OF ANALYSIS

LOT NUMBER: 003M1401
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-20
EXPIRY DATE: 10-June-2020

TEST	SPECIFICATION	RESULT
pH	7.25 – 7.40	7,37
OSMOLALITY	257 - 273	263
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocysts 120h	≥ 80%	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED