

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 001M0201  
PRODUCT NAME: V-GRAD 80%  
REFERENCE: V-GRD-80  
EXPIRY DATE: 31.07.2021

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.48
OSMOLALITY	310-330 mOsm/kg	311
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED