

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 003K4K401  
PRODUCT NAME: V-GRAD Kit  
REFERENCE: V-GRD-KIT  
EXPIRY DATE: 30.09.2021

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7.20-7.60	7.28
pH (V-GRAD 80)	7.20-7.60	7.34
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	329
OSMOLALITY (V-GRAD 80)	310-330 mOsm/kg	324
ENDOTOXIN	< 0.5 EU/mL	PASSED
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED