

CERTIFICATE OF ANALYSIS

LOT NUMBER: 003M0202
PRODUCT NAME: V-GRAD 80%
REFERENCE: V-GRD-80
EXPIRY DATE: 30.09.2021

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.42
OSMOLALITY	310-330 mOsm/kg	318
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED