

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 003M0601  
PRODUCT NAME: V-SPERM FREEZE  
REFERENCE: V-SFM-12  
EXPIRY DATE: 30.09.2021

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.37
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED