

CERTIFICATE OF ANALYSIS

LOT NUMBER: 003M3001
PRODUCT NAME: V-OIL 100 P
REFERENCE: V-OIL-P100
EXPIRY DATE: 30.09.2021

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0.1 mEq/kg	PASSED
ENDOTOXIN	< 0.1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED