

CERTIFICATE OF ANALYSIS

LOT NUMBER: 003M3103
PRODUCT NAME: V-OIL 500 P CE
REFERENCE: V-OIL-P500
EXPIRY DATE: 30.09.2021

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0.1 mEq/kg	PASSED
ENDOTOXIN	< 0.1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
MEA 1-cell to expanded blastocyst in 120 h	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED