



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 003M3202
PRODUCT NAME: V-OIL 100 G
REFERENCE: V-OIL-G100
EXPIRY DATE: 30.09.2021

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0.1 mEq/kg	PASSED
ENDOTOXIN	< 0.1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
MEA 1-cell to expanded blastocyst in 120 h	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED