

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 011M03P01  
PRODUCT NAME: V-GRAD 40%  
REFERENCE: V-GRD43\_100P  
EXPIRY DATE: 31.05.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.38
OSMOLALITY	310 - 340 mOsm/kg	321
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED