



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 011M3301  
PRODUCT NAME: V-OIL 500 G  
REFERENCE: V-OIL-G500  
EXPIRY DATE: 31.05.2022

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0.1 mEq/kg	PASSED
ENDOTOXIN	< 0.1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED