



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 101K4K401
PRODUCT NAME: V-GRAD Kit
REFERENCE: V-GRDK4_12
EXPIRY DATE: 31.07.2022

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7.20-7.60	7.38
pH (V-GRAD 80)	7.20-7.60	7.43
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	333
OSMOLALITY (V-GRAD 80)	310-330 mOsm/kg	327
ENDOTOXIN	< 0.5 EU/mL	PASSED
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED