

CERTIFICATE OF ANALYSIS

LOT NUMBER: 101M01P01
PRODUCT NAME: V-GRAD 100%
REFERENCE: V-GRD00_100P
EXPIRY DATE: 31.07.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7,35
OSMOLALITY	300-330 mOsm/kg	317
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED