



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 101M03P01  
PRODUCT NAME: V-GRAD 40%  
REFERENCE: V-GRD43\_100P  
EXPIRY DATE: 31.07.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.39
OSMOLALITY	310 - 340 mOsm/kg	328
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED