

CERTIFICATE OF ANALYSIS

LOT NUMBER: 101M33C01
PRODUCT NAME: V-GRAD 40%
REFERENCE: V-GRD43_100G
EXPIRY DATE: 31.07.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.39
OSMOLALITY	310 - 340 mOsm/kg	328
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED