

CERTIFICATE OF ANALYSIS

LOT NUMBER: 101M61C01
PRODUCT NAME: V-SPERM FREEZE
REFERENCE: V-SFM-K2_12G
EXPIRY DATE: 31.07.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7,4
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED