

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 102M3202  
PRODUCT NAME: V-OIL 100 G CE  
REFERENCE: V-OIL-G100  
EXPIRY DATE: 31.08.2022

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0.1 mEq/kg	PASSED
ENDOTOXIN	< 0.1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED