

CERTIFICATE OF ANALYSIS

LOT NUMBER: 103M02P01
PRODUCT NAME: V-GRAD 80%
REFERENCE: V-GRD83_100P
EXPIRY DATE: 30.09.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.31
OSMOLALITY	310-340 mOsm/kg	331
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED