

CERTIFICATE OF ANALYSIS

LOT NUMBER: 103M3201
PRODUCT NAME: V-OIL 100 G
REFERENCE: V-OIL-G100
EXPIRY DATE: 30.09.2022

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0.1 mEq/kg	PASSED
ENDOTOXIN	< 0.1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
MEA 1-cell to expanded blastocyst in 120 h	≥ 80 %	PASSED