



VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 104M03P01
PRODUCT NAME: V-GRAD 40%
REFERENCE: V-GRD43_100P
EXPIRY DATE: 31.10.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.39
OSMOLALITY	310 - 340 mOsm/kg	327
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Irena Sherameti ON 03.05.2021

Irena Sherameti