



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 104M61C01
PRODUCT NAME: V-SPERM FREEZE
REFERENCE: V-SFM-K2_12G
EXPIRY DATE: 31.10.2022

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7.50
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED