



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18403VD35, 18405VD50, 18407VD20, 18409VD30, 18404VD40, 18406VD75, 18408VD27

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-135, V-DEN-150, V-DEN-200, V-DEN-300, V-DEN-140, V-DEN-175, V-DEN-275

EXPIRY DATE: 06.09.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED