



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18434VD35, 18435VD40, 18436VD50, 18437VD75, 18438VD20, 18439VD27

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-135, V-DEN-140, V-DEN-150, V-DEN-175, V-DEN-200, V-DEN-275

EXPIRY DATE: 01.10.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

Dr. Ina Schramm

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 13.11.2018
