



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 18518HO20, 18519HO30, 18520HO35, 18521IN302, 18523IN303, 18524IN353

PRODUCT NAME: V-ICSIPET

REFERENCE: V-HOLD-20, V-HOLD-30, V-HOLD-35, V-INJ-S2-30, V-INJ-S3-30, V-INJ-S3-35

EXPIRY DATE: 07.11.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

*i. A. Anna Schrametz*