



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18535IN353, 18536HO20, 18537IN203, 18538IN303, 18539HO30, 18540HO35

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-35, V-HOLD-20, V-INJ-S3-20, V-INJ-S3-30, V-HOLD-30, V-HOLD-35

EXPIRY DATE: 19.11.2021

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

i.A. Irene Sherametz