



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 18554GP23  
PRODUCT NAME: V-GLASSPET 9 (230)  
REFERENCE: V-GPT\_230  
EXPIRY DATE: 24.09.2023

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

*i. A. Irene Shevmetz*

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 18.12.2018

---