



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 185611359
PRODUCT NAME: V-MARKER-PEN
REFERENCE: V-PEN_DC
EXPIRY DATE:

| TEST | SPECIFICATION | RESULT |
|---|---------------|--------|
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |

i.A. Jena Shermet