



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 18DC1029

PRODUCT NAME: V-PURE

REFERENCE: V-PW-2000

EXPIRY DATE: 31.10.2019

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0,25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

i.A. Inema Sherameth

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 21.11.2018
