

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19001GP15
PRODUCT NAME: V-GLASSPET 5.3/4 (150)
REFERENCE: V-GPT_150
EXPIRY DATE: 24.11.2024

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED