

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19002GP23  
PRODUCT NAME: V-GLASSPET 9 (230)  
REFERENCE: V-GPT\_230  
EXPIRY DATE: 24.10.2024

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED