

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19012VD25, 19013VD40, 19014VD75, 19015VD30, 19016VD60

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-125, V-DEN-140, V-DEN-175, V-DEN-300, V-DEN-600

EXPIRY DATE: 25.11.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED