

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19017HO30, 19018HO35, 19019IN303, 19020IN353

PRODUCT NAME: V-ICSIPET

REFERENCE: V-HOLD-30, V-HOLD-35, V-INJ-S3-30, V-INJ-S3-35

EXPIRY DATE: 25.11.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED