

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19035HO30, 19036HO35, 19032IN203, 19033IN303, 19034IN353

PRODUCT NAME: V-ICSIPET

REFERENCE: V-HOLD-30, V-HOLD-35, V-INJ-S3-20, V-INJ-S3-30, V-INJ-S3-35

EXPIRY DATE: 09.12.2022

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED