

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19051VD35, 19052VD40, 19053VD75, 19054VD20, 19055VD27, 19056VD30

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-135, V-DEN-140, V-DEN-175, V-DEN-200, V-DEN-275, V-DEN-300

EXPIRY DATE: 06.01.2023

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED