

CERTIFICATE OF ANALYSIS

LOT NUMBER: 20077HO00, 19064HO30, 19065HO35, 20076IN003, 19066IN203, 19068IN353

PRODUCT NAME: V-ICSIPET

REFERENCE: V-HOLD-00, V-HOLD-30, V-HOLD-35, V-INJ-S3-00, V-INJ-S3-20, V-INJ-S3-35

EXPIRY DATE: 20.01.2023

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED