



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19593IN203, 19594IN303, 19595HO20, 19596HO30, 19597HO35

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-20, V-INJ-S3-30, V-HOLD-20, V-HOLD-30, V-HOLD-35

EXPIRY DATE: 07.01.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED