



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19638HO30, 19639HO35, 19640IN203, 19641IN303, 19649IN353, 19652IN302

PRODUCT NAME: V-ICSIPET

REFERENCE: V-HOLD-30, V-HOLD-35, V-INJ-S3-20, V-INJ-S3-30, V-INJ-S3-35, V-INJ-S2-30,

EXPIRY DATE: 05.02.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED