

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19668VD35, 19669VD40, 19670VD75, 19671VD27, 19673VD60

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-135, V-DEN-140, V-DEN-175, V-DEN-275, V-DEN-600

EXPIRY DATE: 25.02.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED