



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19760IN203, 19761IN303, 19763HO25, 19759HO30

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-20, V-INJ-S3-30, V-HOLD-25, V-HOLD-30

EXPIRY DATE: 06.05.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED