

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19787GP15NP, 19788GP15, 19789GP23NP

PRODUCT NAME: V-GLASSPET

REFERENCE: V-GPT_150NP, V-GPT_150, V-GPT_230NP

EXPIRY DATE: 07.02.2024

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED