

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19842VD25, 19804VD35, 19805VD40, 19806VD50, 19807VD75, 19808VD20, 19809VD27,
19810VD30, 19843VD60

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-125, V-DEN-135, V-DEN-140, V-DEN-150, V-DEN-175, V-DEN-200, V-DEN-275, V-DEN-300,
V-DEN-600

EXPIRY DATE: 01.07.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED