

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19792IN303, 19815IN203, 19816HO30, 19837HO35

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-30, V-INJ-S3-20, V-HOLD-30, V-HOLD-35

EXPIRY DATE: 17.06.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED