

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19845IN353, 19853HO25, 19854HO30

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-35, V-HOLD-25, V-HOLD-30

EXPIRY DATE: 22.07.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED