



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19867VD35, 19868VD40, 19869VD50, 19870VD75, 19871VD20, 19872VD27, 19873VD30

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-135, V-DEN-140, V-DEN-150, V-DEN-175, V-DEN-200, V-DEN-275, V-DEN-300

EXPIRY DATE: 19.08.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED