

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19968IN003, 19935IN353, 19967HO00, 19938HO25, 19936HO30, 19937HO35

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-0, V-INJ-S3-35, V-HOLD-0, V-HOLD-25, V-HOLD-30, V-HOLD-35

EXPIRY DATE: 23.09.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED