

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19949VD20, 19950VD27

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-200, V-DEN-275

EXPIRY DATE: 14.10.2022

| TEST                                      | SPECIFICATION | RESULT |
|---|---------------|--------|
| GAMMA STERILIZATION                       | 25 - 45 kGy   | PASSED |
| ENDOTOXIN                                 | < 0,05 EU/mL  | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 %        | PASSED |