

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 19970VD35, 19971VD40, 19972VD75, 19973VD20, 19974VD27, 19975VD30

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-135, V-DEN-140, V-DEN-175, V-DEN-200, V-DEN-275, V-DEN-300

EXPIRY DATE: 28.10.2022

| TEST                                      | SPECIFICATION | RESULT |
|---|---------------|--------|
| GAMMA STERILIZATION                       | 25 - 45 kGy   | PASSED |
| ENDOTOXIN                                 | < 0,05 EU/mL  | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 %        | PASSED |