

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19976IN353, 19977HO30, 19978HO35

PRODUCT NAME: V-ICSIPET

REFERENCE: V-INJ-S3-35, V-HOLD-30, V-HOLD-35

EXPIRY DATE: 04.11.2022

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED